

Single-layer columns S series (S15,S25)

1. Product Description

Double-layer columns SXK series are designed to connect medium pressure purification systems. Column tube matching adapters can be adapted to a variety of medium pressure purification systems. Double-layer columns SXK series are equipped with Packing reservoir, which is convenient for media filling.

2. Product Information

2.1 Component of the Column		
No.	Accessories Name	
1	Column tube	
2	Column end piece	
3	Glass tube	
4	Column plungers	
5	Bottom plungers	
6	Adjuster knob	
7	Adapter	




Figure 2.1 Component of Column

4

13

6

2.2 Component of Accessories at both ends		
No.	Accessories Name	
8	Stop plugs	
9	Connector for 1/16 tubbings	
10	tubbings	
11	Top cover	
12	Outer rod	
13	Bottom cover	
14	Inner rod	
15	O-ring	
16	Fliter-holder	
17	Filter	

10 6 11 12 4





Figure 2.2 Component of Adapter


Figure 2.3 Component of Bottom plungers



3. Usage method

Figure 2.4 Inner rod

14





3.1 dismantle and installation of Accessories

3.1.1 dismantle and installation of Adapter

1) Remove the adapter

- a. Unscrew the adjuster knob.See step 1 in figure 3.1.
- b. Hold the column end piece, unscrew the top cover.See step 2 in figure 3.1.
- c. Take out the adapter. See step 3 in figure 3.1.

2) Install the adapter

a. Tilt the adapter properly, put it into the column tube.

Note: If there is liquid in the column tube, avoid air bubbles between column plungers and the liquid level when placing the adapter.

- b. Tighten the top cover.
- c. Press and hold the black snap to adjust the adapter position downward.
- d. Tighten the adjuster knob after adjustingh the proper position.

3.1.2 dismantle and installation of Bottom plungers

1) Remove the bottom plungers

The operation steps are the same as 3.1.1.

2) Install the bottom plungers

- a. Place the bottom plungers into the column tube vertically. See step 1 in figure 3.2.
- b. Tighten the bottom cover.See step 2 in figure 3.2.
- c. Tighten the adjuster knob.See step 3 in figure 3.3.

3.1.3 Fix Inner rod

This step is performed after the adapter and the bottom plungers are disassembled to ensure that inner rod structure is fixed.

1) Unscrew the adjuster knob and take down.

2) Take out the inner rod, tighten the screw structure between inner rod and plunger. See figure 3.3.

3) Install the inner rod, tighten the adjuster knob.

3.2 Connect the purifier

Unscrea the stop plugs, connect the connector for 1/16 tubbings to the instrument.



Figure 3.1 Dismantle of Adapter



Figure 3.2 Installation of Bottom plungers



Figure 3.3 Fix Inner rod

4. Troubleshooting

Problem	Probable Cause	Solution
Back pressure too high	Filters are clogged	Clean or replace filters.
	Pipeline is folded and the flow path is	Adjust twisted tubbings.
	not smooth.	
Leakage	Column end piece is leakaged	Tighten the connector between top cover or bottom cover and column end piece.
	Adjuster knob is leakaged	Refer to step 3.1.3 to tighten the inner rod and column plungers.